

User's Manual
for Single Patient Use Only
Blood Glucose Meter



ACCU-CHEK®

Contents

Important Safety Information	3
Chapter 1: Your New System	9
Chapter 2: The Accu-Chek FastClix Lancing Device	19
Chapter 3: Blood Glucose Tests.....	23
Chapter 4: Control Tests.....	37
Chapter 5: Meter Settings.....	43
Chapter 6: Review Your Data.....	63
Chapter 7: Wireless Communication and Meter Pairing.....	81
Chapter 8: Meter and Lancing Device Cleaning and Disinfecting.....	95
Chapter 9: Meter Maintenance and Troubleshooting	103
Chapter 10: Technical Information	113
Chapter 11: License and Warranty	121
Index	125

The Accu-Chek Guide System

The Accu-Chek Guide Blood Glucose Monitoring System is comprised of the Accu-Chek Guide meter and Accu-Chek Guide test strips.

The Accu-Chek Guide Blood Glucose Monitoring System is intended to quantitatively measure glucose in fresh capillary whole blood from the fingertip, palm, and upper arm as an aid in monitoring the effectiveness of glucose control.

The Accu-Chek Guide Blood Glucose Monitoring System is intended for in vitro diagnostic single-patient use by people with diabetes.

The Accu-Chek Guide Blood Glucose Monitoring System is intended to be used by a single person and should not be shared.

This system is not for use in diagnosis or screening of diabetes mellitus, nor for neonatal use.

Alternative site testing should be done only during steady-state times (when glucose is not changing rapidly).

Accu-Chek Guide control solutions are for use with the Accu-Chek Guide Blood Glucose Monitoring System to check that the meters and test strips are working together properly and that the test is performing correctly.

The system includes:

Accu-Chek Guide meter with batteries, Accu-Chek Guide test strips,* Accu-Chek Guide control solutions,* Accu-Chek FastClix lancing device,* Accu-Chek FastClix lancet drums.*

*Some items may not be included in the kit. They are a separate purchase.

WARNING

Choking hazard. Small parts. Keep away from children under the age of 3 years.

NOTE

- The term “blood glucose” is used when referring to “blood sugar.”
- Sample data screens are shown throughout the manual. Your data will differ.

Important Safety Information

Limitations

- Do not use the meter at high hematocrit levels above 65 % or low hematocrit levels below 10 %.
- Not for use in diagnosis or screening of diabetes mellitus.
- Not for neonatal use.
- Abnormally high concentrations of ascorbic acid (vitamin C) resulting in blood concentrations in excess of 5 mg/dL may cause inaccurate results. If you are not sure please check with your doctor.
- Do not use the meter system to measure blood glucose in people who are experiencing cardiovascular collapse (severe shock) or decreased peripheral blood flow.
- Do not use this system during xylose absorption test.
- Not for use on critically ill patients, patients in shock, dehydrated patients, or hyperosmolar patients.
- This system has not been tested at altitudes higher than 10,150 feet.

Important Safety Information

Need Help?

For questions, contact the Accu-Chek Customer Care Service Center toll-free at 1-800-858-8072. Hours of operation are Monday through Friday between 8:00 am and 8:00 pm eastern standard time (EST). We offer assistance in many languages. You can also visit accu-chek.com for diabetes management tools and product demonstrations.

Please complete the warranty card and mail it, so you receive the best customer service possible and product update news. Or register your warranty online at accu-chek.com.

About Testing Yourself or Others

WARNING

- DO NOT CHANGE YOUR THERAPY BASED ON A TEST RESULT THAT DOES NOT MATCH HOW YOU FEEL OR IF YOU BELIEVE THAT YOUR TEST RESULT COULD BE INCORRECT.
- It is always a good idea to have a back-up testing method available. Failure to test could cause a delay in therapy decisions and lead to a serious medical condition. Examples of back-up testing methods include a back-up meter or testing by a laboratory. Ask your healthcare professional or pharmacist about other possible back-up methods.
- If your blood glucose result does not match how you feel and you have followed the instructions in this User's Manual, follow your healthcare professional's instructions, or contact your healthcare professional.

Important Safety Information

WARNING

- During normal testing, any blood glucose meter or lancing device may come in contact with blood. All parts of the kit are considered biohazardous and can potentially transmit infectious diseases from bloodborne pathogens, even after you have performed cleaning and disinfecting.^{1,2}
- The meter and lancing device should never be used by more than one person. Do not share the meter and lancing device with anyone, including family members, due to the risk of infection from bloodborne pathogens.^{1,2} Do not use on multiple patients!
- Cleaning and disinfecting the meter and lancing device destroys most, but not necessarily all, bloodborne pathogens.³
- If the meter is being operated by a second person who is providing testing assistance to the user, the meter and lancing device should be cleaned and disinfected prior to use by the second person.
- Disinfect the meter and lancing device before allowing anyone else to handle them. Do not allow anyone else to test with the meter or lancing device.
- It is important to keep the meter and lancing device clean and disinfected. For instructions on how to clean and disinfect the meter and lancing device, see the chapter **Meter and Lancing Device Cleaning and Disinfecting**.
- Wash hands thoroughly before and after handling the meter, lancing device, or test strips.

Important Safety Information

NOTE

- Perform a control test when you open a new test strip box or if you think that a test result is incorrect. Performing a control test lets you know that the meter and test strips are working properly.
- Refer to the test strip and control solution package inserts for additional health-related information.
- Blood glucose and BG are interchangeable and mean the same thing.

Special Information for Caregivers

- Consult your healthcare professional to determine if it is appropriate for your child to be taught how to use the meter system or any other medical products.
- Some people with diabetes do not experience symptoms of low blood glucose (hypoglycemia). Others, such as children or people who are unconscious or have certain disabilities, may not be able to communicate their symptoms to caregivers. For these reasons, do not change any therapy without first talking to a healthcare professional.

Important Safety Information

Before You Start Testing

WARNING

- Carefully read and follow the instructions in the User's Manual and package inserts for the test strips and control solutions. It is very important to follow the instructions in order to avoid an incorrect test result that leads to improper therapy.
- Inspect the test strip container before using the test strips for the first time. If you see any damage to the container cap or if anything prevents the cap from closing properly, do not use the test strips. Contact the Accu-Chek Customer Care Service Center. Damaged test strips can cause inaccurate results, which could lead to improper therapy.
- The meter, test strips, and control solution are only for use outside the body (in vitro). Do not eat the test strips. Do not swallow or inject the control solution or use the control solution for any purpose other than testing the Accu-Chek Guide system.

NOTE

- The meter may prompt you to choose a language the first time you turn it on.
- Check the time and date on the meter before you begin testing. Adjust the time and date if necessary.

The Accu-Chek Guide Meter



Front View



Back View



Side View

1. Display

Shows results, messages, and test results stored in memory.

2. Back Button

Returns to a previous display or field.

3. Up Arrow and Down Arrow Buttons

Press to move between menu options or to increase or decrease numbers.

4. Power/Set/OK Button

Turns meter on or off and sets options.

5. Test Strip Slot with Light

Insert test strip here.

6. Battery Door

Flip open to replace batteries.

7. Micro USB Port


Transfers data from the meter to a computer (PC).

8. Test Strip Ejector

Press to remove test strip.



(for example)



Test Strip



13



14

9. Test Strip Container*

10. Metallic End

Insert this end into meter.

11. Yellow Edge

Touch blood drop or control solution here.

12. Control Solution Bottle*

13. Batteries

14. USB Cable*

Connects the meter to a PC.

*Some items may not be included in the kit.
They are a separate purchase.

The Accu-Chek FastClix Lancing Device



- 1. Cap**
For fingertip testing.
- 2. Comfort Dial with Depth Selection**
Select penetration depth.
- 3. Penetration Depth Indicator**
Points to depth setting.

- 4. Lever**
Advance to a new lancet.
- 5. Lancet Counter Window**
Shows number of available lancets.
- 6. Release Button**
Press to prime and prick.
- 7. Lancet Drum**
Contains 6 lancets.



NOTE

Some items may be sold separately.


1 Your New System

Button Functions


Here are the functions of the back, arrow, and Power/Set/OK buttons on the meter. These functions are used throughout this manual. See the chapter **Meter Settings** for specific instructions on setting up the meter.

Button	Function
 Back Button	Return to the previous display. Return to the previous field.
 Up Arrow and Down Arrow Buttons	Navigate up and down in a menu. Increase or decrease a number.
 Power/Set/OK Button	Press briefly to turn the meter on. Press and hold to turn the meter off. Press to select an option. Press to move to the next field or display. Press to save an option. With the meter off, press and hold to check the meter display. Display Check


Meter Menus



- ❶ Title of display or menu
- ❷ Menu options
- ❸ Scroll down
- ❹ Scroll up



Display	Description
	There are more menu options listed beneath the last option. Press  on the meter to view the options.
	There are more menu options listed above the first option. Press  on the meter to view the options.
 	There are more menu options listed above and below the options. Press  or  on the meter to view the options.

Display	Description
	Highlighted option (Test) Press OK to enter the Test menu.
	Highlighted field (Hour) Press ▼ or ▲ to increase or decrease the hour. Press OK to set the hour and move to the minutes field.

Symbols

Here is a list of the symbols on the meter display.

Symbols	Description
	Above target range
	After meal
	Bedtime
	Before meal
	Below target range
	Blood glucose test
	Checkmark / Control test OK / Selected option or setting
	Control bottle
	Control test not OK
	Edit
	Error
	Fasting

Symbols	Description
	Flight mode
	Help
	Low battery
	My data
	No comment
	Other
	Overall
	Settings
	Test reminder
	Warning
	Within target range

Set the Language

The meter may prompt you to choose a language the first time you turn it on.


1



Turn the meter on by briefly pressing **OK**. **Language** appears.

Press **▼** or **▲** to highlight the language.

2



Press **OK** to set the desired language and return to **Main Menu**.

NOTE

If you select the wrong language and cannot correct it, contact the Accu-Chek Customer Care Service Center at 1-800-858-8072.

Using the Accu-Chek FastClix Lancing Device



- The black cap is for fingertip testing only.
- Remember that the opening where the lancet comes out is not in the center of the cap.
- Remove the cap by pulling it straight off. Do not twist the cap.



- A new lancet drum is dark gray with a white end.



- A used lancet drum has a visible red stripe on the white end.
- You cannot reuse a used lancet drum.
- Do not remove the lancet drum until you have used all 6 lancets.



- The penetration depth indicator on the comfort dial shows the current depth setting. The higher the number, the deeper the penetration. The best depth setting is the lowest number that lets you get enough blood for a test. Try different depth settings to find the one that is right for you.
- For soft skin, we suggest a depth setting of 2. For thick skin, try a higher depth setting.


2

The Accu-Chek FastClix Lancing Device

Inserting a Lancet Drum

You must first load the lancet drum into the lancing device to get it ready for use.


1



Remove the cap.

Insert a new lancet drum, white end first, until it clicks firmly into place.

2



Slide the cap on until it stops by aligning the notch on the cap with the notch on the lancing device.

3



You are now ready to use the first lancet.

The lancet counter shows a number 6, meaning you have 6 new lancets remaining.

NOTE


Once the lancet drum is inserted into the lancing device, do not remove the lancet drum until completely used. The lancet drum cannot be reused once it has been removed from the lancing device.

WARNING

You must not insert the lancet drum into the lancing device and simultaneously press the release button or hold the lancing device with the release button resting on a surface such as a table top. This could release a lancet and inadvertently cause injury.

Using the Accu-Chek FastClix Lancing Device

1



Adjust the lancet depth by turning the comfort dial.

2

Obtain a blood drop. See the chapter **Blood Glucose Tests**.

3



Advance to the next lancet by sliding the lever forward and back all the way.

The number in the lancet counter decreases by 1.

The lancet counter window shows the number of remaining lancets.


NOTE

- ALWAYS use a new, sterile lancet each time you test to avoid infection.
- For safety reasons, once you advance to a new lancet, you cannot go back to a used lancet.
- NEVER reuse a lancet.
- NEVER share your lancing device with anyone.

Changing the Lancet Drum

When you have used the sixth and last lancet, change the lancet drum.

1



Remove the cap.

Hold the lancet drum between your thumb and index finger and pull it straight out.

A red stripe is visible on the white part, indicating that the lancet drum has been used.

2



Throw the old lancet drum away.

Always discard according to local regulations.

3




Insert a new lancet drum, white end first, until it clicks firmly into place.

Slide the cap back on until it stops by aligning the notch on the cap with the notch on the lancing device.

WARNING

- The meter and lancing device should never be used by more than one person. Do not share the meter and lancing device with anyone, including family members, due to the risk of infection from bloodborne pathogens.^{1,2} Do not use on multiple patients!
- Blood glucose results can be displayed in either mg/dL or mmol/L. The back label of the meter shows the unit of measurement. If the meter shows the wrong unit, contact the Accu-Chek Customer Care Service Center at 1-800-858-8072. The correct unit of measure in the US is mg/dL. If you do not know which unit of measurement is correct for you, contact your healthcare professional. Using the wrong unit of measurement may cause misinterpretation of your actual blood glucose level and may lead to improper therapy.



3 Blood Glucose Tests

Using the Accu-Chek Guide System

- The Accu-Chek Guide test strips are for testing fresh capillary whole blood.
- Use only Accu-Chek Guide test strips.
- Use the test strip immediately after removing it from the test strip container.
- Do not apply blood or control solution to the test strip before inserting it into the meter. If you applied blood or control solution before inserting the test strip into the meter, retest with a new test strip. If a result appears before applying blood or control solution, do not act on that result.
- Close the test strip container tightly immediately after removing a test strip. Moisture can damage the test strips and produce incorrect results.
- Store the unused test strips in their original container with the cap tightly closed.
- Discard the test strips if they are past the **Use By** date printed on the test strip container. If the **Use By** date is missing or cannot be read, do not use the test strips. Contact the Accu-Chek Customer Care Service Center at 1-800-858-8072.
- Refer to the test strip package insert for test strip storage and system operating conditions.
- Do not remove test strips from the test strip container and put them into another container, such as a plastic bag, pocket, purse, wallet, etc.
- Do not reuse test strips. Once control solution or blood has been applied to a test strip, discard it. If a retest is necessary, use a new test strip.
- Perform a control test every time you open a new test strip box.

WARNING

To prevent inaccurate results:

- DO NOT expose test strips to heat, moisture, or humidity. Temperatures outside the required range, as well as moisture and humidity, can damage the test strips and lead to inaccurate results.
- DO NOT bend, cut, or alter the test strips.
- DO NOT get dirt, food, or other material on the test strip.

Performing a Blood Glucose Test with Blood from Your Fingertip

Refer to the **Important Safety Information** section at the beginning of this manual.

NOTE

- Before you perform your first blood glucose test, set up the meter correctly.
- You need the meter, a test strip, and a lancing device with a lancet drum loaded to perform a blood test.
- A blood glucose test cannot be performed while the meter is connected to a PC with a USB cable.
- There are 2 ways to start a blood glucose test.
 - Insert a test strip into the meter.
 - Turn the meter on by briefly pressing . Select **Test** > .

1



Wash your hands with warm soapy water and dry thoroughly. Dirty or wet hands could affect test results.

2



Adjust the comfort dial.

3



Check the **Use By** date on the test strip container.


Do not use test strips past the **Use By** date.

4




Remove a test strip from the test strip container. Close the cap tightly.

5



Insert the metallic end of the test strip into the meter.



The meter turns on. **Preparing to test** appears.

6



When **Apply drop** appears, obtain a blood drop.


7



Press the lancing device firmly against the side of your fingertip.


Press the release button all the way down to prick your finger.

8




Gently squeeze your finger to assist the blood flow. This helps you get a blood drop.


9



Touch the **yellow edge** of the test strip to the blood drop. Remove your finger from the test strip when **Analyzing** appears. Do not put blood on top of the test strip.



10



The test result appears on the display.

You have the option of adding a comment to the test result by pressing **OK** OR proceed to Step 13 to complete the test.


3 Blood Glucose Tests

11



Add Comment appears. Press to highlight a comment. Press to set the comment for the test result. See the **Adding Comments to Blood Glucose Results** section of this chapter for details.

12



The final result appears. Press or to set the comment and return to **Main Menu**. Or to change the comment, press to select the comment.

Press to return to **Add Comment**.


13



Press here to eject test strip

Remove and discard the used test strip by pulling the test strip out of the meter or by pushing the test strip ejector on the side of the meter.

14



Advance to the next lancet.

15



Wash hands thoroughly with soap and water.

Blood Glucose Warnings

If your blood glucose result is outside the measurement range of the meter, a warning is displayed. Press **OK** to acknowledge the LO or HI warning, OR the meter automatically moves to the LO or HI result display.

⚠ WARNING LO Result Retest BG. Contact your healthcare professional.	9:38am 12/11/15 LO Add Comment	Blood glucose may be lower than the measurement range of the system. See the Unusual Blood Glucose Results section of this chapter.
⚠ WARNING HI Result Consider checking BG, ketones, and insulin.	9:38am 12/11/15 HI Add Comment	Blood glucose may be higher than the measurement range of the system. See the Unusual Blood Glucose Results section of this chapter.

3 | Blood Glucose Tests

Adding Comments to Blood Glucose Results

NOTE

Analyzing your blood glucose results stored in the meter is an effective way for you and your healthcare professional to determine how well you are controlling your diabetes. This analysis is a valuable tool for making improvements to your diabetes management. Use care when adding comments to blood glucose results. Incorrect comments can cause inaccurate patterns to be detected by the meter if Patterns is **On**.

Overview

It is very important to have the correct time and date set in the meter. Having the correct time and date setting helps ensure accurate interpretation of information by you and your healthcare professional.

- You may add comments to blood glucose results to help you and your healthcare professional analyze patterns detected by the meter (see the **Patterns** section in the chapter **Meter Settings** for details).
- If Patterns is set to **On**, once a pattern is detected you may NOT change the comment attached to a blood glucose result (see the **Patterns** section in the chapter **Meter Settings** for details).
- Adding a comment saves the comment and the symbol with the blood glucose results.


Here is a list of comment symbols that can be added to a blood glucose result.

Symbols	Name	Description
	Before meal	If Patterns is On , select Before breakfast, lunch, dinner, or snack (see the following page for adding comments with Patterns On).
	After meal	If Patterns is On , select After breakfast, lunch, dinner, or snack (see the following page for adding comments with Patterns On).
	Fasting	Select Fasting for no caloric intake for at least 8 hours.*
	Bedtime	
	Other comment	You can use this comment to mark an event such as an AST result or exercise.
	No entry	1. You do not want to add a comment. 2. You want to remove a comment for the current blood glucose result.

* American Diabetes Association: Standards of Medical Care in Diabetes-2013.

3 Blood Glucose Tests

1



After performing a blood glucose test, the test result is displayed on the screen with **Add Comment** highlighted. Press **OK** to add a comment.

2




The **Add Comment** menu appears. Press **OK** to highlight the desired comment. (The example here is **Before meal**.) Press **OK** to select the comment.



If Patterns feature is On:

If **Patterns** is **On** and you select **Before meal** or **After meal**, press **OK** to select a specific meal (**Breakfast**, **Lunch**, **Dinner**, or **Snack**). Press **OK** to set the selected meal for the test result.

3



The final result appears with the comment symbol. Press **OK** or **Back** to set the comment and return to **Main Menu**.

To change the comment, press **OK** to select the comment. Press **OK** to return to **Add Comment**.

Performing a Blood Glucose Test with Blood from Your Palm or Upper Arm (Alternate Site Testing)

WARNING

- Do not use Alternate Site Testing to calibrate a continuous glucose monitoring system or to make insulin dosing calculations.
- Alternate site testing should be done only during steady-state times (when glucose is not changing rapidly).

You have the option of obtaining a blood sample from other sites on your body besides the fingertip. Alternate sites include the palm and upper arm.

Blood obtained from the fingertip can be used at any time to test blood glucose.

If blood from an alternate site is used, there are certain times when testing is not appropriate. This is because your blood glucose level changes faster in your fingertip than in the alternate sites. These differences may cause you to misinterpret your actual blood glucose level, leading to improper therapy and potential adverse health effects.

Read the next section before you try testing from alternate sites.

You may perform a palm or upper arm test

- immediately before a meal.
- while fasting.

You may NOT perform a palm or upper arm test

- up to 2 hours following a meal, when blood glucose values can rise quickly.
- after injecting bolus insulin, when blood glucose values can decrease rapidly.
- after exercise.
- if you are sick.
- if you think your blood glucose is low (hypoglycemia).
- if you sometimes do not notice when your blood glucose is low.

If you are interested in AST, talk to your healthcare professional first.

To obtain an AST cap and detailed AST instructions, contact the Accu-Chek Customer Care Service Center at 1-800-858-8072.

3 | Blood Glucose Tests

Unusual Blood Glucose Results

If your blood glucose result does not match how you feel, follow these steps:

1. Perform a control test. See the chapter **Control Tests**.
2. Repeat the blood glucose test.

If your blood glucose result still does not match how you feel, follow your healthcare professional's instructions or call your healthcare professional immediately.

NOTE

Always follow your healthcare professional's instructions. For example, if your healthcare professional has advised you to immediately treat a low blood glucose result (such as by eating something), then do that first.

WARNING

- Do not change your treatment because of one blood glucose result.
- NEVER ignore symptoms of low or high blood glucose.

Symptoms of Low or High Blood Glucose

WARNING

The meter is designed to provide a numerical value for blood glucose in the range of 20–600 mg/dL.

Being aware of the symptoms of low or high blood glucose can help you understand your test results and decide what to do if they seem unusual.

Low blood glucose (hypoglycemia): Symptoms of hypoglycemia may include, but are not limited to, anxiety, shakiness, sweating, headache, increased hunger, dizziness, pale skin color, sudden change in mood or irritability, fatigue, difficulty concentrating, clumsiness, palpitations, and/or confusion.

High blood glucose (hyperglycemia): Symptoms of hyperglycemia may include, but are not limited to, increased thirst, frequent urination, blurred vision, drowsiness, and/or unexplained weight loss.

If you are experiencing any of these symptoms, or other unusual symptoms, test your blood glucose from the fingertip. If your blood glucose result is displayed as LO or HI, follow your healthcare professional's instructions or contact your healthcare professional immediately. If your blood glucose result does not match how you feel, follow the steps in **Unusual Blood Glucose Results**.

Comparing Your Meter Result to a Laboratory Result

A common question is how the blood glucose results on the meter compare to the laboratory results. Your blood glucose can change quickly, especially after eating, taking medication, or physical activity. If you test yourself in the morning, then go to your healthcare professional's office for a blood glucose test, your test results will probably not match, even if you are fasting. This is typically not a problem with the meter, it just means that time has elapsed and your blood glucose level has changed.

Although you always apply fresh capillary whole blood to the test strip, the system has been calibrated to deliver plasma-like values for easier comparison to laboratory results.

If you want to compare your meter result to the laboratory result, **you must be fasting**. Take the meter to your healthcare professional's office and test yourself by fingerstick within 5 minutes of having blood drawn from your arm by a healthcare professional. Keep in mind that the laboratory uses different technology than the meter and that blood glucose meters for self-testing generally read somewhat lower than the laboratory result.

When to Perform a Control Test

Performing a control test lets you know the meter and test strips are working properly. You should perform a control test when:

- you open a new test strip box.
- you left the test strip container open.
- you think the test strips are damaged.
- you want to check the meter and test strips.
- the test strips were stored in extreme temperatures, humidity, or both.
- you dropped the meter.
- your test result does not match how you feel.
- you want to check if you are performing the test correctly.

About the Control Solutions

- Use only Accu-Chek Guide control solutions.
- Close the control solution bottle tightly after use.
- Write the date you open the control solution bottle on the bottle label. The control solution must be discarded 3 months from the date the control solution bottle was opened (discard date) or on the **Use By** date on the bottle label, whichever comes first.
- Do not use control solution that is past the **Use By** or discard date.
- Refer to the control solution package insert for control solution storage conditions.
- The meter automatically recognizes the difference between the control solution and blood.
- The control results are not displayed in memory.
- The control solution can stain fabric. Remove stains by washing with soap and water.
- Control solution is available for purchase. To order the control solution, talk to your pharmacist or visit accu-chek.com to order online.

4 Control Tests

Performing a Control Test

You need the meter, a test strip, and control solution Level 1 or Level 2.

1




Check the **Use By** date on the test strip container. Do not use test strips past the **Use By** date.

2




Remove a test strip from the test strip container. Close the cap tightly.

3



Insert the metallic end of the test strip into the meter. Place the meter on a flat surface.



The meter turns on. **Preparing to test** appears.



Apply drop appears.

4




Select the control solution to test. You will enter the level later in the test.

5




Remove the bottle cap. Wipe the tip of the bottle with a tissue. Squeeze the bottle until a tiny drop forms at the tip.

6



Touch the drop to the **yellow edge** of the test strip. Do not put control solution on top of the test strip.



Analyzing appears when there is enough control solution in the test strip.

7



Control Result and the control bottle symbol appear.

Press or to select the control level you tested. If you do not select a level, the control result is saved without a control level.

Press .

8



or



Within range and appear if the control result is within range.

Out of range and appear if the control result is out of range.

9



Wipe the tip of the bottle with a tissue. Cap the bottle tightly.

Remove and discard the used test strip.

NOTE

- Most people just test the Level 1 control. If you wish, you can also test a Level 2 control. A set of Level 1 and Level 2 control solutions is available for purchase.
- The ranges for Level 1 and Level 2 control solutions are printed on the test strip container label.
- The meter turns off 90 seconds after a successful test or 15 seconds after the test strip is removed, provided no other action is taken.

Understanding Out-of-Range Control Results

WARNING

The control range applies only to a control result. It only indicates that the test strips and meter are working properly. Do not use a control result to interpret blood glucose results.

If the control result is out of range, do not use the meter until you solve the problem. Check this list to help solve the problem.

Troubleshooting Checks	Action
Were the test strips or control solutions expired?	Discard the test strips or control solution if either is past the Use By date. If the control solution was opened more than 3 months ago, discard it. Repeat the control test with an unexpired test strip and an unexpired control solution.
Did you wipe the tip of the control solution bottle before use?	Wipe the tip of the bottle with a tissue. Repeat the control test with a new test strip and a fresh drop of control solution.
Were the caps on the test strip container and the control solution bottle always closed tightly?	Replace the test strips or control solution if you think either was uncapped for some time. Repeat the control test.
Was the test strip used immediately after it was removed from the test strip container?	Repeat the control test with a new test strip and a fresh drop of control solution.
Were the test strips and control solutions stored in a cool, dry place?	Repeat the control test with a properly stored test strip or control solution.
Did you follow the directions?	Read the chapter Control Tests and repeat the control test.
Did you choose the correct control solution level, either 1 or 2, when you performed the test?	If you chose the wrong control solution level, you can still compare the control result to the range printed on the test strip container.
Are you still unsure of the problem?	Contact the Accu-Chek Customer Care Service Center at 1-800-858-8072.

Overview


You can adjust the following settings in the meter for your personal preferences. Refer to the sections later in this chapter for details and how to set the options.

Setting	Options	Function
Time/Date	Time / Date	Set the time and date.
Beeper	On / Off	Select On or Off.
Wireless	Select wireless communication settings. See the chapter Wireless Communication and Meter Pairing .	
Target Ranges	Off / Single Range / 2 Ranges	Select the blood glucose target range appropriate for you. NOTE Consult your healthcare professional for the appropriate target range for you. Off – no target range arrow symbols appear with blood glucose result. Patterns is Off (see the Patterns section of this chapter for details).
	70–160 mg/dL (pre-set target range)	Single Range – blood glucose results are marked as above, within, or below range based on the single target range set in the meter.
	Before Meal Range 70–110 mg/dL (pre-set target range)	2 Ranges – set Before Meal and After Meal ranges. Blood glucose results are marked as above, within, or below range based on the 2 target ranges (Before Meal and After Meal) set in the meter.
	After Meal Range 70–160 mg/dL (pre-set target range)	

Setting	Options	Function
Patterns	On / Off	On – a pattern is detected when 2 below-target or 3 above-target results with the same comment are detected within a 7-day period.
Reminders	On / Off / Edit time	On – set up to 4 reminders per day to remind you to test
Post Meal	On / Off / Edit time	On – reminds you to perform an after meal blood glucose test.
Last Result	On / Off	Select whether the previous blood glucose result (within the past 24 hours) appears with the current blood glucose result. On – the previous blood glucose result appears with the current blood glucose result. Off – only the current blood glucose result appears.
Language		Select the language for the meter.

Time/Date

1



Turn the meter on by briefly pressing **OK**. From **Main Menu**, press **▼** to highlight **Settings**. Press **OK**.

2



Time/Date is highlighted. Press **OK**.

3



Press **▼** or **▲** to adjust each field. Press **OK** to set and proceed to the next field. Set **am** or **pm** if necessary.


Press **OK** to save and return to the previous menu.

Beeper

The beeper prompts you:

- when a test strip is inserted.
- to apply blood or control solution to the test strip.
- when enough blood or control solution is drawn into the test strip.
- when the blood glucose or control test is complete.
- when the meter is turned on.
- when a button is pressed.
- when it is time to perform a test (if Reminders or Post Meal reminders are **On**).
- when the batteries are inserted.
- when there are no stored blood glucose results or there is an invalid record in the logbook.
- when there are no errors in the error log.
- if an error occurred (even if the beeper is off, it still beeps for an error).

1




Turn the meter on by briefly pressing **OK**. From **Main Menu**, press **▼** to highlight **Settings**. Press **OK**.

2



Press **▼** to highlight **Beeper**. Press **OK**.

3



Press **▼** or **▲** to highlight **On** or **Off**. Press **OK** to move **✓** to the option.

Press **OK** to set the option and return to the previous menu.

Target Ranges

Your healthcare professional can tell you what blood glucose range is appropriate for you. It is very important to stay within your target range.

Target Ranges can be set from a lower limit of 60–100 mg/dL to an upper limit of 101–300 mg/dL.

Options	Function
Off	Arrow symbols for above, within, or below target ranges do not appear with the blood glucose results.
Single Range	Set lower limit and upper limit for the target range. You will be prompted to turn on Patterns if desired (see the Patterns section of this chapter for details).
2 Ranges	Set lower limits and upper limits for Before meal and After meal target ranges. You must mark your blood glucose results with a comment for the meter to detect above, within, or below Before Meal or After Meal test results (see the Adding Comments to Blood Glucose Results section in the chapter Blood Glucose Tests for details). You will be prompted to turn on Patterns if desired (see the Patterns section of this chapter for details).


When Target Ranges is **On**, the following symbols appear with blood glucose results.

Symbol	Meaning
	The blood glucose result is below the target range.
	The blood glucose result is within the target range.
	The blood glucose result is above the target range.

WARNING

This function is no substitute for hypoglycemia training by your healthcare professional.

1



Turn the meter on by briefly pressing **OK**. From **Main Menu**, press **▼** to highlight **Settings**. Press **OK**.

2



Press **▼** to highlight **More Options**. Press **OK**.

3



Target Ranges is highlighted. Press **OK**.

4



Press **▼** or **▲** to highlight **Off**, **Single Range**, or **2 Ranges**. Press **OK** to select an option (the example here is **2 Ranges**).


5



Press **▼** or **▲** to adjust the lower limit of the **Before Meal** target range. Press **OK** to set and move to the next field.


Continue to set the upper limit for the **Before Meal** target range and the lower and upper limits for the **After Meal** target range. Press **OK**.

6



The ✓ appears next to the option you selected.

Press  to return to the previous menu.



The meter may prompt you to turn on Patterns.

Press  or  to highlight **Yes** or **No**.

Press  to set the option and return to the previous menu (see the **Patterns** section of this chapter for details).

NOTE


Analyzing your blood glucose results stored in the meter is an effective way for you and your healthcare professional to determine how well you are controlling your diabetes. This analysis is a valuable tool for making improvements to your diabetes management. Use care when adding comments to blood glucose results. Incorrect comments can cause inaccurate patterns to be detected by the meter if Patterns is **On**.

Patterns

A Pattern is detected by the meter when 2 below-target (Low Pattern) or 3 above-target (High Pattern) test results with the same comment are detected within a 7-day period.

- The meter does NOT detect a pattern for the “Other” comment added to blood glucose results.
- It is very important to have the correct time and date set in the meter. Having the correct time and date setting helps ensure accurate interpretation of information by you and your healthcare professional.
- Only blood glucose results marked with comments will be included in Patterns. If LO or HI test results are marked with comments, the results become part of Patterns (see the chapter **Blood Glucose Tests** for more details).
- Target Ranges must be set in the meter to use Patterns. If Target Ranges are not set, the meter prompts you to set them.

1



Turn the meter on by briefly pressing **OK**. From **Main Menu**, press **▼** to highlight **Settings**. Press **OK**.

2




Press **▼** to highlight **More Options**. Press **OK**.

3




Press **▼** to highlight **Patterns**. Press **OK**.

4




Press **▼** to highlight **On**. Press **OK** to move **✓** to the option. Press **OK**.



This message appears if Target Ranges is Off:


Press **▼** to highlight **Yes**. Press **OK**.

(To turn Patterns **Off**, select **No**. Press **OK** to return to **Patterns**.)




Target Ranges appears on the display. (see the **Target Ranges** section of this chapter for details on setting target ranges).

If Patterns is **On** and a new pattern is detected with a blood glucose result, a message appears on the display.



Press  to select **Details** to view the blood glucose results that make up that pattern.

or



Press  to highlight **View Later**.

Press  to return to the previous screen.

Reminders

You can set up to 4 general test Reminders per day to remind you to test. A series of beeps sound and  is displayed for Reminders set in the meter.


Reminders:

- turn off by inserting a test strip or pressing any button.
- are postponed until the next test reminder if a test was performed within 15 minutes of a test reminder.
- do not appear/beep if the meter is on at the test reminder time (including when the meter is connected to a USB cable).
- do not appear/beep if the meter is connected and communicating to a Continua Certified® manager.
- do not beep if the meter's beeper is set to off.
- do not appear/beep if the batteries need to be replaced.

Set Reminders

- Reminder times are pre-set in the meter for 8:00 am, 12:00 pm, 6:00 pm, and 10:00 pm. You may change reminder times by following the instructions below.
- If a general test Reminder is set for the same time as a Post Meal reminder, the Post Meal reminder will appear/beep instead of the general Reminder (see the **Post Meal Reminders** section of this chapter for details).

1



Turn the meter on by briefly pressing **OK**. From **Main Menu**, press **▼** to highlight **Settings**. Press **OK**.

2



Press **▼** to highlight **More Options**. Press **OK**.

3




Press **▼** or **▲** to highlight **Reminders**. Press **OK**.

4




Pre-set reminder times appear on the display. Press to highlight a reminder time. Press .

5




The pre-set reminder time appears. Press or to select **On** or **Off**. Press to move to the option.




To change the Reminder time shown:

Press to highlight **Edit time**. Press .



Press or to adjust each field. Press to move to the next field. Press to return to the previous menu.

6



A ✓ indicates the reminder is set for the time shown and automatically set to **On**. To turn the reminder off, press to highlight **Off**.

Press to return to **Reminders**.

7



The reminder time appears.

Press to continue to set additional reminders or press to return to **More Options**.

Post Meal Reminders

Post Meal reminders can be set to remind you to test again later when you add a Before Meal comment to a blood glucose result. When a reminder occurs, a series of beeps sound and  is displayed.


Post Meal reminders:

- turn off by inserting a test strip or pressing any button.
- are postponed until the next test reminder if a test was performed within 15 minutes of a test reminder.
- do not appear/beep if the meter is on at the test reminder time (including when the meter is connected to a USB cable).
- do not appear/beep if the meter is connected and communicating to a Continua Certified® manager.
- do not beep if the meter's beeper is set to off.
- do not appear/beep if the batteries need to be replaced.

Set Post Meal Reminders

- Adding a Before Meal comment to a blood glucose result sets a Post Meal reminder in the meter.
- Marking blood glucose results with a Post Meal comment provides more information about your test results to help you and your healthcare professional in the management of your diabetes.
- Talk to your healthcare professional to determine your Post Meal test time.
- Select 1 hour, 1.5 hours, or 2 hours for Post Meal reminders to occur.

1



Turn the meter on by briefly pressing **OK**. From **Main Menu**, press **▼** to highlight **Settings**. Press **OK**.

2



Press **▼** to highlight **More Options**. Press **OK**.

3



Press **▼** to highlight **Post Meal**. Press **OK**.

5

Meter Settings

Post Meal Reminders

4

Post Meal	
✓	On
	Off
✎	Edit time
?	Help

Press or to highlight **On**.
Press to move ✓ to the option.

5

Post Meal	
✓	On
	Off
✎	Edit time
?	Help

Press or to highlight **Edit time**.
Press .

6

Post Meal	
Remind after:	

	1 hour
✓	1.5 hours
	2 hours

Press or to highlight **1 hour, 1.5 hours, or 2 hours** after a meal for **Post Meal** reminders to occur.


Press to move ✓ to the option.

Press to set and return to the previous menu.

Last Result


Select whether the previous blood glucose result appears with the current blood glucose result.

Test results older than 24 hours do not appear.



Off – only the current blood glucose result ❶ appears.

On – the previous blood glucose result ❷ appears with the current blood glucose result.



Turn the meter on by briefly pressing **OK**. From **Main Menu**, press **▼** to highlight **Settings**. Press **OK**.

5

Meter Settings

Last Result

2




Press  to highlight **More Options**. Press .

3



Press  to highlight **Last Result**. Press .

4




Press  or  to highlight **On** or **Off**. Press  to move  to the option.

Press  to set the option and return to the previous menu.

Language

Choose the language that appears on the meter.

1



Turn the meter on by briefly pressing **OK**. From **Main Menu**, press **▼** to highlight **Settings**. Press **OK**.

2



Press **▼** to highlight **More Options**. Press **OK**.

3



Press **▼** to highlight **Language**. Press **OK**.

4



Press **▼** or **▲** to highlight the desired language. Press **OK** to move **✓** to the option.

Press **OK** to set the language and return to the previous menu.

Overview

- Blood glucose results are stored from the newest to the oldest.
- The meter automatically stores up to 720 blood glucose results in memory with the time and date of the test and any test result comments.
- Once 720 blood glucose results are in memory, adding a new blood glucose result deletes the oldest blood glucose result.
- Only test results that have been marked with a fasting, before meal, after meal, or bedtime comment are included in the average for that comment.
- All test results are included in the overall 7, 14, 30, and 90-day averages regardless of what comment is added.
- Control results are stored in memory but cannot be reviewed on the meter. To view stored control results, transfer them to a compatible software application.
- Control results are not included in the averages or blood glucose reports.
- Once 32 control results are in memory, adding a new control result deletes the oldest control result.

WARNING

Do not change your therapy based on an individual test result in memory. Talk to your healthcare professional before changing therapy based on test results in memory.


6

Review Your Data

Logbook


Logbook

1




Turn the meter on by briefly pressing **OK**. From **Main Menu**, press **▼** to highlight **My Data**. Press **OK**.

2



Logbook is highlighted. Press **OK**.

3




Press **▼** or **▲** to scroll through **Logbook**.

The most recent test result ①.

The 2nd most recent test result ②.

- 4 To view details about a test result, press  or  to highlight the test result. Press . Test result details shown below only appear if **Target Ranges** is **On** or comments were added to a test result.

1



The most recent test result.

2



The 2nd most recent test result.


6

Review Your Data

Averages


Averages

1



Turn the meter on by briefly pressing **OK**. From **Main Menu**, press **▼** to highlight **My Data**. Press **OK**.

2



Press **▼** to highlight **Averages**. Press **OK**.

3



Press **▼** to highlight a category (the example here is **Overall**). Press **OK**.

4



Press **▼** to highlight a time period (the example here is **90 days**). Press **OK**.

5



Press **⬅** to return to the previous menu if you want to review a different time period OR press **▼** or **▲** to move through different averages.

Target Percent (%)

Target Percent (%) allows you to view the percentage of your Overall, Before meal, After meal, Fasting, and Bedtime blood glucose results that are above, within, or below your target ranges.


- Target % results can be viewed for 7, 14, 30, or 90-day time periods.
- Target Ranges must be set in the meter to review Target % results.

1



Turn the meter on by briefly pressing **OK**. From **Main Menu**, press **OK** to highlight **My Data**. Press **OK**.

2



Press **OK** to highlight **Target %**. Press **OK**.

3



Press **OK** to highlight a category (the example here is **Before meal**). Press **OK** to select the option.

6

Review Your Data

Target Percent (%)

4

Target %
🍏 Before meal
7 days
14 days
30 days
90 days

Press  to highlight a time period (the example here is **90 days**). Press .

5

Target %
🍏 Before meal
90 days
⬆ Above
⬆ Within
⬇ Below
Tests:
720

The **Target %** appears (for the **Before meal** example). The number of total tests included in the Target % appears at the bottom of the display.

Press  to return to the previous menu.

NOTE

Analyzing your blood glucose results stored in the meter is an effective way for you and your healthcare professional to determine how well you are controlling your diabetes. This analysis is a valuable tool for making improvements to your diabetes management. Use care when adding comments to blood glucose results. Incorrect comments can cause inaccurate patterns to be detected by the meter if Patterns is **On**.

Low/High Data

Your healthcare professional can tell you what blood glucose range is appropriate for you. It is very important to stay within your target range.

- Target Ranges must be set in the meter to track Low/High Data test results (see the **Target Ranges** section in the chapter **Meter Settings** for details).
- Low BG or High BG Data includes only results that fall above or below the target ranges set in the meter.
- Low BG or High BG Data is tracked in the meter for 30 days.

6

Review Your Data


Low/High Data

You can select Low BG and High BG results for Overall, Before meal, After meal, Fasting, or Bedtime blood glucose results.

Symbol	Name	Description
	Overall	Includes high and low blood glucose results based on Target Ranges set in the meter.
	Before meal	You may view low or high test results marked with a Before Meal comment for Overall, Before breakfast, Before lunch, Before dinner, and Before snack blood glucose results.*
	After meal	You may view low or high test results marked with an After Meal comment for Overall, After breakfast, After lunch, After dinner, and After snack blood glucose results.*
	Fasting	Includes high or low fasting blood glucose results marked as Fasting in comments.
	Bedtime	Includes high or low bedtime blood glucose results marked as Bedtime in comments.

*Test results for Before and After specific meals are only available if Patterns is set to **On**.

1




Turn the meter on by briefly pressing **OK**. From **Main Menu**, press **▼** to highlight **My Data**. Press **OK**.

2



Press **▼** to highlight **Low/High Data**. Press **OK**.

3



If Target Ranges is set to On:

Press **▼** to select **Low** or **High BG Data** (the example here is **High BG Data**). Press **OK**.

or



If Target Ranges have NOT been On within the last 30 days:

This message appears on the meter (to turn on Target Ranges see the **Target Ranges** section in the chapter **Meter Settings** for details).

4




Press to highlight a category (the example here is **Before meal**). Press .



If test results with detailed meal comments are saved in the Logbook:

The meter may prompt you to select detailed categories to view. Press to highlight a category (the example here is **before Breakfast**). Press .

5



The selected data appears (the example here is **High BG Data**). Press to scroll through the test results.

Press to return to the previous menu.


NOTE

Analyzing your blood glucose results stored in the meter is an effective way for you and your healthcare professional to determine how well you are controlling your diabetes. This analysis is a valuable tool for making improvements to your diabetes management. Use care when adding comments to blood glucose results. Incorrect comments can cause inaccurate patterns to be detected by the meter if Patterns is **On**.


Patterns

- Patterns displays only active Low Patterns or High Patterns based on comments added to blood glucose results within the last 7 days.
- A Pattern is generated when 2 below-target or 3 above-target test results with the same comment are detected within a 7-day period.

Patterns may be viewed on the meter in 3 ways:



or



or



when a Patterns option is displayed at the bottom of **Main Menu**.

from **My Data** on **Main Menu**.

if a **New pattern detected** message appears on the display when performing a blood glucose test.

6


Review Your Data

Patterns

Patterns detected by the meter may be displayed on the Main Menu as:


Patterns	High and low patterns have been detected	Patterns may include the following blood glucose results marked with comments: Before breakfast, After breakfast, Before lunch, After lunch, Before dinner, After dinner, Before snack, After snack, Fasting, or Bedtime (see the Adding Comments to Blood Glucose Results section in the chapter Blood Glucose Tests).
High Pattern(s)	One or more high patterns have been detected	
Low Pattern(s)	One or more low patterns have been detected	
No Patterns	No active pattern based on results from last 7 days	
Blank	Patterns feature is set to Off	

1



Turn the meter on by briefly pressing **OK**. From **Main Menu**, press **▼** to highlight the **Patterns** option at the bottom of the screen (the example here is **Patterns**). Press **OK**.

2



Press **▼** or **▲** to highlight **Low Patterns** or **High Patterns** (the example here is **Low Patterns**). Press **OK** to select the option.

3



Press **▼** to highlight a category to review (the example here is **Before breakfast**). Press **OK**.

4



Press **▼** to scroll through test results.
Press **▲** to return to the previous menu.

6


Review Your Data

Data Transfer Using Wireless

Data Transfer Using Wireless

This feature allows you to transfer data wirelessly from your meter to another device.

1




Turn the meter on by briefly pressing **OK**. From **Main Menu**, press **▼** to highlight **My Data**. Press **OK**.

2




Press **▼** to highlight **Data Transfer**. Press **OK**.

3



Press **▼** to select **Wireless**. Press **OK**.

4



or






If a device has not been paired with your meter:

The meter prompts you to pair a device to the meter (see the **First-Time Pairing** section in the chapter **Wireless Communication and Meter Pairing** for details).

If the meter has already been paired with at least one device:

Press  or  to choose a wireless device for data transfer.

5



The meter transfers the data to the device.

6


Review Your Data

Data Transfer Using USB Cable

Data Transfer Using USB Cable


This feature allows you to transfer data from your meter to special software for diabetes management on a PC using a USB cable.

1



Plug the small end of the USB cable into the meter.

2



Plug the large end of the USB cable into a USB port on the PC.
If the meter is off, it turns on.

3

Start the software for data analysis and initiate a data transfer.

4

Data Transfer



Connecting

Data Transfer

Transferring
data

Data Transfer



Complete

The meter transfers the data to the software.

Overview

You can wirelessly and automatically synchronize your diabetes information with a mobile device. The process of creating a connection between the meter and the other device is called pairing. You need an application on your mobile device that can accept the meter's data.

Setting	Options	Function
Flight Mode	On / Off	Select whether wireless communication is available. On – wireless communication is not available. Off – wireless communication is available.
Default Device	List of devices paired with the meter	If more than 1 device is paired, select the device with which the Auto-Send and Sync Time features will communicate.
Auto-Send	On / Off	Select whether data is automatically sent to the default paired device after each test. On – data is automatically sent to the default paired device. Off – data is not automatically sent to the default paired device.
Sync Time	On / Off	Select whether to synchronize the time and date to the default paired device. On – the time and date on the meter synchronize to the time and date on the default paired device. Off – the time and date on the meter do not synchronize to the time and date on the default paired device.
Pairing	Pair Device / Delete Pairing	Select whether to pair a device or to delete a paired device.

First-Time Pairing

The steps below are for first-time pairing.


Up to 5 devices can be paired with the meter at any time.

Each device must be paired with the meter one at a time.


The meter and the device to be paired should be within 1 meter of each other.

1

Prepare the device for pairing.
Refer to the device's
instructions for information
about pairing.


2

Turn the meter on by briefly
pressing **OK**. From **Main Menu**,
press **▼** to highlight **Settings**.
Press **OK**.

3

Press **▼** to highlight **Wireless**.
Press **OK**.

4



Pair Device appears. To pair a device, press to highlight **Yes**. Press .

5



The meter displays its code.

6

The mobile device displays a list of found devices, including the meter. Select the meter from the list. When prompted, enter the code displayed on the meter from Step 5 into the device to be paired.

7




When pairing is complete, and the name of the paired device appear.

Press to return to the previous menu.

Flight Mode

Select whether wireless communication is available or not. When **Flight Mode** is on,  appears in the title bar and wireless communication is not available.

1




Turn the meter on by briefly pressing **OK**. From **Main Menu**, press **▼** to highlight **Settings**. Press **OK**.

2




Press **▼** to highlight **Wireless**. Press **OK**.

3




Flight Mode is highlighted. Press **OK**.

4



Press **▼** or **▲** to highlight **On** or **Off**. Press **OK** to move **✓** to the option.

Press **OK** to set the option and return to the previous menu.




Main Menu with meter in Flight Mode.

Default Device

If more than 1 device is paired, select the default paired device for the **Auto-Send** and **Sync Time** features.

1




Turn the meter on by briefly pressing **OK**. From **Main Menu**, press **▼** to highlight **Settings**. Press **OK**.

2



Press **▼** to highlight **Wireless**. Press **OK**.

3



Press **▼** to highlight **Default Device**. Press **OK**.

4




A list of paired devices appears. Press **▼** to highlight the device you want as the default device. Press **OK** to move **✓** to the option. Press **OK** to set the option and return to the previous menu.

Auto-Send


Select whether data is automatically sent to the default paired device after each test.

1




Turn the meter on by briefly pressing **OK**. From **Main Menu**, press **▼** to highlight **Settings**. Press **OK**.

2




Press **▼** to highlight **Wireless**. Press **OK**.

3



Press **▼** to highlight **Auto-Send**. Press **OK**.

4




Press **▼** or **▲** to highlight **On** or **Off**. Press **OK** to move **✓** to the option.

Press **OK** to set the option and return to the previous menu.

Sync Time

Select whether to synchronize the time and date to the default paired device.

1




Turn the meter on by briefly pressing **OK**. From **Main Menu**, press **OK** to highlight **Settings**. Press **OK**.

2




Press **OK** to highlight **Wireless**. Press **OK**.

3



Press **OK** to highlight **Sync Time**. Press **OK**.

4



Press **OK** or **Left Arrow** to highlight **On** or **Off**. Press **OK** to move **✓** to the option.

Press **OK** to set the option and return to the previous menu.

Pairing Additional Devices


This procedure is to add another paired device (assumes there is already 1 paired device).

You can pair 5 devices.

If you have 5 paired devices, you must delete a pairing before you can add another device.

The device you choose remains the default until you change it.

1




Turn the meter on by briefly pressing **OK**. From **Main Menu**, press **▼** to highlight **Settings**. Press **OK**.

2



Press **▼** to highlight **Wireless**. Press **OK**.

3



Press **▼** to highlight **Pairing**. Press **OK**.

4



Press or to highlight **Pair Device**. Press .

If there are less than 5 paired devices:

Pair Device appears. Proceed to Step 5.



If there are already 5 paired devices:

Maximum paired devices appears. You must delete a paired device before you can proceed to Step 5. Press or to highlight **Yes**. Press .



Press or to highlight the device to delete. Press . Proceed to Step 5.

5

Prepare the other device for pairing. Refer to the other device's instructions for information about pairing.

6



The meter displays its code.

7

The mobile device displays a list of found devices, including the meter. Select the meter from the list. When prompted, enter the code displayed on the meter from Step 6 into the device to be paired.

8

Pair Device



Paired with
Device1234

When pairing is complete, ✓
and the name of the paired
device appear. Press **OK**.

9

Pair Device

Set as default?

Device1234

No

Yes

If there is more than 1 paired device, the meter asks if you want
the new device to be the default device. Press **▼** or **▶** to highlight
No or **Yes**. Press **OK** to set the option and return to **Pairing**.

Press **⏮** to return to the previous menu.


Delete Pairing

This procedure is to delete a paired device (assumes there is at least 1 paired device).

You can pair 5 devices. If you have 5 paired devices, you must delete a pairing before you pair to another device.


The device you choose remains the default until you change it.

1




Turn the meter on by briefly pressing **OK**. From **Main Menu**, press **▼** to highlight **Settings**. Press **OK**.

2



Press **▼** to highlight **Wireless**. Press **OK**.

3



Press **▼** to highlight **Pairing**. Press **OK**.

4



Press or to highlight **Delete Pairing**. Press .

5




Press or to highlight the device to delete. Press to set the option.

6

**If the selected device is not the default device:**

The pairing is deleted. The name of the deleted device and ✓ appear.

**If the selected device is the default device:**

The meter prompts you to select a new default device. If there is only 1 device remaining, it becomes the default device. The list of paired devices appears.



Highlight the new default device and press . The name of the deleted device and ✓ appear, along with the new default device.

What is the difference between cleaning and disinfecting?

Cleaning is the removal of dirt from the meter or lancing device.³

Disinfecting is the removal of most, but not all, disease-causing and other types of microorganisms (bloodborne pathogens) from the meter or lancing device.³

Approved Cleaning and Disinfecting Product

The following product has been approved for cleaning and disinfecting the meter and lancing device:

Super Sani-Cloth (EPA* reg. no. 9480-4)

Super Sani-Cloth can be purchased from Amazon.com, Officedepot.com, and Walmart.com.

- Do not use any other cleaning or disinfecting solutions. Using solutions other than the Super Sani-Cloth could result in damage to the meter and lancing device.
- The effect of using more than one product interchangeably to clean and disinfect the meter and lancing device has not been tested. Always use Super Sani-Cloth to clean and disinfect the meter and lancing device.
- Roche has demonstrated that the product is good for 5-year use, after testing in a total of 260 disinfection cycles (equal to cleaning and disinfecting once per week for 5 years).

*Environmental Protection Agency

8 Meter and Lancing Device Cleaning and Disinfecting

NOTE

For technical assistance or questions on cleaning and disinfecting, contact the Accu-Chek Customer Care Service Center at 1-800-858-8072.

Cleaning and Disinfecting the Meter

WARNING

If the meter is being operated by a second person who is providing testing assistance to the user, the meter and lancing device should be cleaned and disinfected prior to use by the second person.

To clean and disinfect without damaging the meter, follow these procedures carefully.

When to Clean and Disinfect the Meter

- Clean the meter to remove visible dirt or other material prior to disinfecting.
- Clean and disinfect the meter at least once per week and when blood is present on the surface of the meter.
- Clean and disinfect the meter before allowing anyone else to handle the meter. Do not allow anyone else to use the meter on themselves for testing purposes.

NOTE

Using cleaning and disinfecting products could result in damage to the meter. If you notice any of the following signs of deterioration after cleaning and disinfecting your meter, stop using your meter and contact the Accu-Chek Customer Care Service Center at 1-800-858-8072: residue around buttons, clouding of display, button malfunction, out-of-range control results.

What to Clean and Disinfect

The following parts of the meter should be cleaned and disinfected:

- The area around slots and openings (do not get any moisture in slots or openings)
- The meter display
- The entire meter surface

How to Clean and Disinfect the Meter

WARNING

Failure to follow these instructions will damage the meter and stop it from working properly.

- DO NOT clean or disinfect the meter while performing a blood glucose or control test.
- DO NOT get any moisture in slots or openings.
- DO NOT spray anything onto the meter.
- DO NOT immerse the meter in liquid.
- Always use the same product for both cleaning and disinfecting.




1



Wash hands thoroughly with soap and water.

2



Turn the meter off and wipe the entire meter surface with a Super Sani-Cloth. Carefully wipe around the test strip slot and other openings.

Make sure that no liquid enters any slot or opening.

3

A separate Super Sani-Cloth should be used for cleaning and disinfection. For disinfecting the meter, get a new cloth and repeat step 2, making sure the surface stays wet for 2 minutes.

Make sure that no solution is seen in any slot or opening.

8 Meter and Lancing Device Cleaning and Disinfecting

4



Wash hands thoroughly with soap and water.

Cleaning and Disinfecting the Lancing Device

To clean and disinfect without damaging the lancing device, follow these procedures carefully.

When to Clean and Disinfect the Lancing Device

- Clean the lancing device to remove visible dirt or other material prior to disinfecting.
- Clean and disinfect the lancing device at least once per week to remove visible dirt or other material for safe handling.
- Clean and disinfect the lancing device before allowing anyone else to handle the lancing device, for instance, if you have someone assisting you. Do not allow anyone else to use the lancing device.

NOTE

- Do not throw away the cap after each use. Use the approved cleaning and disinfecting product on the cap.
- Always remove the lancet drum before cleaning or disinfecting the lancing device.
- Using cleaning and disinfecting products could result in damage to the lancing device. If you notice any of the following signs of deterioration after cleaning and disinfecting your lancing device, stop using your lancing device and contact the Accu-Chek Customer Care Service Center at 1-800-858-8072: residue around buttons, difficulty in priming the device, difficulty in inserting the lancet drum.
- You might observe a slight discoloration of the lancing device after multiple cleaning and disinfecting cycles. This does not affect the functionality of the lancing device.

What to Clean and Disinfect

The following parts of the lancing device should be cleaned and disinfected:

- The entire lancing device surface
- The cap

8 Meter and Lancing Device Cleaning and Disinfecting

How to Clean and Disinfect the Lancing Device

WARNING

Failure to follow these instructions may damage the lancing device and stop it from working properly.


- DO NOT get any moisture into any openings.
- Always use the same product for both cleaning and disinfecting.

1



Wash hands thoroughly with soap and water.

2



Wipe the entire surface of the lancing device and the inside of the cap with a Super Sani-Cloth.

3

A separate Super Sani-Cloth should be used for cleaning and disinfection. For disinfecting the lancing device, use a new cloth and repeat step 2 making sure the surface stays wet for 2 minutes.

4



Wash hands thoroughly with soap and water.

8 Meter and Lancing Device Cleaning and Disinfecting

Meter Maintenance

The meter automatically tests its own systems every time you turn it on and lets you know if something is wrong. See the **Error Messages** section of this chapter.

If you have problems with the meter or think the results are not accurate, perform a control test with an unexpired test strip and control solution. If the control result is not within the acceptable range, contact the Accu-Chek Customer Care Service Center at 1-800-858-8072.

Changing the Batteries



NOTE

- The meter uses two 3-volt lithium batteries, coin cell type CR2032. This type of battery can be found in many stores. It is a good idea to have spare batteries available.
- Always replace both batteries at the same time and with the same brand.
- The logbook data is saved when you replace the batteries.

Error Messages

 WARNING

- Never make therapy decisions based on an error message.
- If you have any concerns or see any other error message, contact the Accu-Chek Customer Care Service Center at 1-800-858-8072.



The meter will not turn on or the display is blank.

- Batteries are dead.

Insert new batteries.

- Display is damaged. / Meter is defective.

Contact the Accu-Chek Customer Care Service Center at 1-800-858-8072.

- Extreme temperatures.

Move the meter to a more temperate area.



 **Test Not Allowed**
Remove strip
or USB cable.

The meter is connected to a PC and a test cannot be performed.

EITHER remove the USB cable and perform a test OR remove the test strip and start a data transfer.



 **Connection Lost**
Reconnect USB
cable and retry.

The connection between the meter and PC was lost.

Disconnect and reconnect the USB cable and retry the connection. Contact the Accu-Chek Customer Care Service Center at 1-800-858-8072 if the connection is lost again.

No Connection

Connect USB cable to continue.

There is no connection between the meter and PC.

Connect the USB cable and retry the connection.

Transfer Failed

Ensure paired device is within range and turned on.

Blood glucose results were not transferred to a paired device.

Make sure the paired device is within range of the meter and turned on.

Connection Failed

Check PC or USB cable.

Data could not be transferred from the meter to the PC.

Check the PC or USB cable.

Transfer Not Allowed

Wireless communication not allowed in flight mode.

Data cannot be sent to a paired device because the meter is in Flight Mode.

Retry the data transfer when the meter is not in Flight Mode.

Results Excluded

One or more results may be excluded from averages.

One or more blood glucose results are excluded from the selected averages because the results are invalid or outside the system measurement range.

Results Excluded

One or more results may be excluded from target % data.

One or more blood glucose results are excluded from the selected target % data because the test results are invalid.

 **Results
Excluded**

One or more
results may be
excluded from
low/high data.

One or more blood glucose results are excluded from the selected low/high data because the test results are invalid.

 Invalid Date

Check the date.

The date entered is not valid.
Enter the correct date.

 **Flight Mode
Active**

Wireless is off.
Selected setting
was saved.

A meter setting was changed while in Flight Mode.

The setting change will not take effect until Flight Mode is turned off.

 **Flight Mode
Active**

Wireless is off.
Pairing is not
allowed.

Pairing to a device cannot be performed while in Flight Mode.
Retry pairing when the meter is not in Flight Mode.

 **Auto-Send
Not Complete**

Transfer still
pending. Ensure
paired device
is within range
and turned on.

The blood glucose result has not been sent to the default paired device. The transfer is still pending.

Place meter and paired device closer together.

 Pairing Failed

Retry pairing.

The meter was unable to pair with a device.

Retry the pairing.

⊗ Strip Error

Reinsert or
replace with
new strip.

(E-1)

The test strip may be damaged, not properly inserted, or was previously used.

Remove and reinsert the test strip or replace it if damaged or previously used.

⊗ Test Error

Retest with
new strip.

(E-3)

A meter or test strip error has occurred.

Repeat the blood glucose test.

If a second E-3 code appears, perform a control test with the control solution and a new test strip.

- If the control result is within the acceptable range, review the proper testing procedure and repeat the blood glucose test with a new test strip.
- If the control result is not within the acceptable range, see the **Understanding Out-of-Range Control Results** section in the chapter **Control Tests**.

In rare cases, an E-3 error code may indicate that your blood glucose is extremely high and above the system's reading range. If you do not feel like your blood glucose is extremely high, repeat the blood glucose test. See the **Unusual Blood Glucose Results** section in the chapter **Blood Glucose Tests**. If the E-3 code still appears for your blood glucose test, **contact your healthcare professional immediately**.

⊗ Drop Too Small

Retest with new strip.

(E-4)

Not enough blood or control solution was drawn into the test strip for measurement or was applied after the test had started.

Discard the test strip and repeat the blood glucose or control test.

⊗ Drop Applied Early

Retest with new strip.

(E-6)

Blood or control solution was applied to the test strip before **Apply drop** appeared.

Discard the test strip and repeat the blood glucose or control test.

⊗ Electronic Error

Remove batteries, press any button, reinsert batteries.

(E-7)

An electronic error occurred. Remove the batteries, press and hold the Power/Set/OK button for at least 2 seconds, and reinsert the batteries. Perform a blood glucose or control test.

⊗ Temperature Error

Move meter to correct temperature.

(E-8)

The temperature is above or below the proper range for the system.

Refer to the test strip package insert for system operating conditions. Move to an area with the appropriate conditions and repeat the blood glucose or control test. Do not artificially heat or cool the meter.

⊗ Replace Batteries

Replace batteries now.

(E-9)

The batteries are out of power.

Change the batteries now. If the message reappears after the batteries have been replaced, remove the batteries, press and hold the Power/Set/OK button for at least 2 seconds, then reinsert the batteries.

⊗ Time/Date
Error

Correct
time/date
if necessary.

(E-10)

The time and date setting may be incorrect.

Make sure the time and date are correct and adjust, if necessary.

⊗ Test Error

Retest with
new strip.

(E-11)

The test strip may be damaged.
Retest with a new test strip.

⊗ High Ascorbate

Contact your
healthcare
professional.

(E-12)

Your blood sample may contain a high level of ascorbate.
Contact your healthcare professional.

⊗ Strip Error

Reinsert or
replace with
new strip.

(E-13)

Fluid or foreign material may be present in the test strip slot. Remove and reinsert the test strip or repeat the blood glucose or control test with a new test strip. If the problem persists, contact the Accu-Chek Customer Care Service Center at 1-800-858-8072.

⊗ Electronic
Error

Contact Roche.

(E-14)

An electronic error has occurred.

Contact the Accu-Chek Customer Care Service Center at 1-800-858-8072.

⊗

(E-14)

Sync Time

Meter time has
been updated to
match time on
paired device.

The meter time and date have been changed to match the paired device.

Logbook

No stored results

There are no results in the Logbook.

Logbook

--- mg/dL
#100 ---
Invalid result
▼

There is an invalid result in the Logbook.

Average
After meal
7 days

No data for
selection

Tests:

Target %
After meal
7 days

No data for
selection

Tests:

There are no test results in range for the selected data.

Target %

No data available.
Enable target
ranges to track
target data.

Target Ranges is Off and there are no results for the Target % data stored in the meter.

Low/High Data

No data available.
Enable target
ranges to track
low/high BG data.

Target Ranges is Off and there are no results for the low/high data stored in the meter.


⬆ High BG Data

No high BG data
available

⬇ Low BG Data

No low BG data
available


There are no results stored in the meter for the selected data.



Patterns is **On** but there are no active patterns stored in the meter.



Patterns is **Off**.



Blood glucose may be higher than the measurement range of the system.

See the **Unusual Blood Glucose Results** section in the chapter **Blood Glucose Tests**.



Blood glucose may be lower than the measurement range of the system.

See the **Unusual Blood Glucose Results** section in the chapter **Blood Glucose Tests**.

Product Limitations

See the literature packaged with the test strips and control solutions for the latest information on product specifications and limitations.

Specifications	
Blood volume	Refer to the test strip package insert.
Sample type	
Measuring time	
Measuring range	
Test strip storage conditions	
System operating conditions	
Meter storage conditions	Temperature: -13–158 °F
Memory capacity	720 blood glucose results and 32 control results with time and date
Automatic off	90 seconds
Power supply	Two 3-volt lithium batteries (coin cell type CR2032)
Display	LCD
Dimensions	80 × 47 × 20 mm (LWH)
Weight	Approx. 40 g (with batteries)
Construction	Hand-held
Protection class	III
Meter type	The Accu-Chek Guide meter is suitable for continuous operation.
Control solution storage conditions	Refer to the control solution package insert.
Interfaces	USB: micro-B connector; Bluetooth Low Energy; Continua Certified® to a Continua Certified manager.
Radio frequency connectivity	Bluetooth low energy technology operating in the frequency band of 2402 MHz to 2480 MHz with a maximum transmitted power of 0 dBm (1 mW).



10 | Technical Information

Bluetooth® Wireless Technology –The meter uses *Bluetooth* Low Energy wireless technology to communicate and transfer information. *Bluetooth* wireless technology is a form of radio frequency (RF) technology that operates in the unlicensed industrial, scientific and medical band at 2.4 to 2.485 GHz. The RF channel utilized for communication between the meter and other devices, such as a smartphone, is not an open channel. The meter can only communicate with the device that

1. utilizes *Bluetooth* Low Energy technology,
2. it is paired with, and
3. has an application that can accept the meter's data.

This device complies with United States Federal Communication Commission (FCC) standards. The device complies with FCC Part 15 Rules. Operation of the device is subject to the following conditions:

1. This device may not cause harmful interference and
2. must accept any interference received, including interference that may cause undesired operation.

Compliance with these guidelines means that under normal, daily circumstances, the device should not affect the operation of other devices. In addition, the device should operate normally in the presence of other devices.

In the event there is interference from another device, it is recommended that you increase the distance between the meter and that device. You can also turn off the interfering device. In addition, you can turn off *Bluetooth* Low Energy wireless technology on the meter. Changes or modifications to the device not expressly approved by Roche could void the user's authority to operate the device. The device has been tested and found to comply with the limits for a Class B digital device. The device generates, uses, and can radiate radio frequency energy.

Electromagnetic Compatibility – This meter meets the electromagnetic immunity requirements as per ISO 15197. The chosen basis for electrostatic discharge immunity testing was basic standard IEC 61000-4-2. In addition, the meter meets the electromagnetic emissions requirements as per EN 61326. The meter's electromagnetic emission is thus low. Interference from the meter to other electrically-driven equipment is not anticipated.

This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions:

1. This device may not cause harmful interference, and
2. this device must accept any interference received, including interference that may cause undesired operation.

This equipment complies with FCC and IC radiation exposure limits set forth for an uncontrolled environment. This equipment is in direct contact with the body of the user under normal operating conditions. This transmitter must not be co-located or operating in conjunction with any other antenna or transmitter.

NOTE

This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

Changes or modifications not expressly approved by the party responsible for compliance (i.e. the manufacturer) could void the user's authority to operate the equipment.

10 | Technical Information

Performance Analysis – Refer to the test strip package insert.

Test Principle – Refer to the test strip package insert.

Product Safety Information

WARNING

- This meter meets IEC 61010-1, IEC 61010-2-101 and IEC 60950-1 safety standards.
- Strong electromagnetic fields may interfere with the proper operation of the meter. Do not use the meter close to sources of strong electromagnetic radiation.
- To avoid electrostatic discharge, do not use the meter in a very dry environment, especially one in which synthetic materials are present.

Travel Documentation

If you are traveling on a commercial airline, you may be required to provide documentation certifying that this meter meets environmental conditions and test procedures for Airborne Equipment (RTCA DO-160) section 21 Emission of Radio Frequency Energy. Visit accu-chek.com or contact the Accu-Chek Customer Care Service Center at 1-800-858-8072 to obtain a copy of the document.

Discarding the Meter, Test Strips, Lancing Devices, Lancets, and Batteries

WARNING

- Any product coming in contact with blood is considered contaminated (potentially infectious).*
- During normal testing, any blood glucose meter may come in contact with blood.
- Lancing devices may also be considered sharps. Disposal of sharps is regulated by law in many jurisdictions.

Roche is committed to recycling and sustainability. Comply with any laws or ordinances relating to the disposal of sharps and/or contaminated products. Contact your local health department or other appropriate authorities for proper handling and disposal of used meters, used test strips, used lancets, and used batteries.

Consider the following points when discarding used testing materials: Consider recycling the meters and batteries at an appropriate facility. Be aware the meter is potentially hazardous electronics scrap (e-scrap) and should be disposed of accordingly. The batteries are potentially hazardous also and should be discarded accordingly.

Disinfect the meter before recycling or discarding.

*29 CFR 1910.1030 – Bloodborne pathogens

Explanation of Symbols

	Caution, refer to safety-related notes in the instructions for use accompanying this product.
	Biological Risks – used meters carry a risk of infection.
 FCC ID: WX3-126	This device complies with Part 15 of the FCC Rules.
	Global Trade Item Number
	Serial number
	3-volt coin cell type CR2032

Additional Supplies

Test Strips: Accu-Chek Guide test strips

Control Solutions: Accu-Chek Guide control solutions

Lancets: Accu-Chek FastClix 102-ct. lancet drums (17-6 ct. drums)

References

¹ FDA Public Health Notification: “Use of Fingerstick Devices on More than One Person Poses Risk for Transmitting Bloodborne Pathogens: Initial Communication, (2010). Update 11/29/2010.” <http://www.fda.gov/MedicalDevices/Safety/AlertsandNotices/ucm224025.htm>. Accessed June 1, 2016.

² CDC Clinical Reminder: “Use of Fingerstick Devices on More than One Person Poses Risk for Transmitting Bloodborne Pathogens, (2010).” <http://www.cdc.gov/injectionsafety/Fingerstick-DevicesBGM.html>. Accessed June 1, 2016.

³ Healthcare Infection Control Practices Advisory Committee (HICPAC), William A. Rutala, Ph.D., M.P.H., and David J. Weber, M.D., M.P.H. Centers for Disease Control and Prevention, 2008. “Guideline for Disinfection and Sterilization in Healthcare Facilities.” Atlanta. http://www.cdc.gov/hicpac/pdf/guidelines/Disinfection_Nov_2008.pdf. Accessed June 1, 2016.

Warranty

Accu-Chek Guide Meter 30-day Money-back Guarantee for Qualifying Consumers

Roche offers qualifying consumers that purchase an Accu-Chek Guide meter, a 30-day money back guarantee. If you are not fully satisfied with your Accu-Chek Guide meter, contact the Accu-Chek Customer Care Service Center toll-free at 1-800-858-8072 to determine whether you qualify to receive a full refund within 30 days of purchase. If you are covered under Medicare, Medicaid, other federal/state programs, or private insurance you are NOT eligible for this money-back offer. Consumers affected by this exclusion may instead request a different Accu-Chek meter/system. The refund will be limited to the amount paid by you net of any rebates. You must have a copy of the dated itemized purchase receipt and the original packaging to obtain this refund.

Accu-Chek Guide Meter Limited 3-Year Warranty

Roche warrants to the original purchaser of the meter that your Accu-Chek Guide meter will be free from defects in materials and workmanship for three years from the date of purchase. If, during this 3-year period, the meter does not work properly because of a defect in materials or workmanship, Roche will replace it with a new Accu-Chek Guide meter or equivalent product free of charge. The warranty on the replacement meter will expire on the date of the original warranty expiration or 90 days after the shipment of a replacement system, whichever period is longer. The purchaser's exclusive remedy with respect to the Accu-Chek Guide meter shall be replacement.

This warranty does not apply to the performance of an Accu-Chek Guide meter that has been damaged by accident or has been altered, misused, tampered with, or abused in any way. Roche will handle meters that show damage or abuse according to its Non-Warranty Service Policy described on the following page.

THE ABOVE WARRANTY IS EXCLUSIVE OF ALL OTHER WARRANTIES, AND ROCHE MAKES NO OTHER WARRANTIES, EXPRESS OR IMPLIED, INCLUDING WITHOUT LIMITATION, THE IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. IN NO EVENT SHALL ROCHE BE LIABLE TO THE PURCHASER OR ANY OTHER PERSON FOR ANY INCIDENTAL, CONSEQUENTIAL, INDIRECT, SPECIAL, OR PUNITIVE DAMAGES ARISING FROM OR IN ANY WAY CONNECTED WITH THE PURCHASE OR OPERATION OF THE METER OR ITS PARTS. NO WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE, IF ANY IS IMPLIED FROM THE SALE OF THE METER, SHALL EXTEND FOR A LONGER DURATION THAN THREE YEARS FROM THE DATE OF PURCHASE.

Some states do not allow limitations on how long an implied warranty will last or the exclusion of incidental or consequential damages, so the above limitation and exclusion may not apply to you. This warranty gives you specific legal rights, which vary from state to state.

11 | License and Warranty

Non-Warranty Service Policy

Roche Non-Warranty Service Policy applies to meters where the above warranty has not become effective, has become inapplicable, or has expired. Roche will replace, at its option, meters returned to it for a service charge (not to exceed \$35).

Replacement will be with the same or similar product. Replacement meters will be warranted for a period of 90 days from shipment under a limited warranty providing for replacement of parts and labor at no charge.

Warranty and Service Instructions

All requests for return of Accu-Chek Guide meters under the above warranty or service policy must be made to the Accu-Chek Customer Care Service Center. You will be mailed a return authorization label, which must be affixed to your carton for shipping the system to Roche. Cartons received without this label will be returned to you at your expense.

Customers experiencing difficulties should review the troubleshooting information in **Meter Maintenance and Troubleshooting** of this manual. Further inquiries should be directed to the Accu-Chek Customer Care Service Center.

Be sure to fill out and mail the Warranty Card that comes with the Accu-Chek Guide system.

Limited License

WARNING

CAUTION – A RESTRICTED LICENSE LIMITS USE OF THE ACCU-CHEK GUIDE SYSTEM (meter and test strips) IN THE UNITED STATES – READ CAREFULLY THE LIMITATIONS RECITED BELOW.

The Accu-Chek Guide system (meter and test strips) and its use are protected by U.S. Patent Nos. 6,645,368 (expires 22-December-2017); 7,276,146 (expires 4-October-2022); 7,276,147 (expires 4-October-2022); 8,349,168 (expires 4-October-2022); 8,425,759 (expires 4-October-2022); 9,017,543 (expires 4-October-2022); 9,017,544 (expires 4-October-2022); 8,298,401 (expires 4-October-2022); 8,303,801 (expires 4-October-2022); 8,329,026 (expires 4-October-2022); 7,407,811 (expires 9-May-2020); 7,452,457 (expires 2-May-2026); 7,488,601 (expires 1-February-2026); 7,494,816 (expires 29-December-2019); 8,377,707 (expires 17-October-2023); 8,859,293 (expires 17-October-2023); 7,569,126 (expires 28-December-2026); and 8,361,291 (expires 8-October-2024). A license to use the Accu-Chek Guide system is required until the expiration of the last-to-expire patent listed above and is only granted when the Accu-Chek Guide meter is used with the Accu-Chek Guide test strips.

Accu-Chek Guide test strips are specifically manufactured for operation with the Accu-Chek Guide meter. Use of other test strips supplied by another manufacturer may prevent or impair the proper function of the Accu-Chek Guide system.

Using the Accu-Chek Guide system indicates your acceptance of the restricted license to use the Accu-Chek Guide system only with Accu-Chek Guide test strips. If you do not agree to the terms and conditions of the restricted license, you may return, at the place of purchase, the unused Accu-Chek Guide system for a full refund. If you have any questions, please call the Accu-Chek Customer Care Service Center at 1-800-858-8072.

Except where prohibited by statute, all warranties covering the Accu-Chek Guide system are voided by use of the Accu-Chek Guide system with any test strips other than Accu-Chek Guide test strips.

11 License and Warranty

WARNING

A RESTRICTED LICENSE LIMITS USE OF THE ACCU-CHEK FASTCLIX SYSTEM (lancing device and lancet drums) IN THE UNITED STATES. READ CAREFULLY THE LIMITATIONS RECITED BELOW.

The Accu-Chek FastClix system (device and lancet drums) and its use are protected by U.S. Patent Nos. 7,322,998 (expires 3-March-2020); and 7,785,338 (expires 5-January-2026). A license to use the Accu-Chek FastClix system is required until the expiration of the last-to-expire patent listed above and is only granted when Accu-Chek FastClix lancet drums are used with the Accu-Chek FastClix device.

Accu-Chek FastClix lancet drums are high precision components that are produced to the close tolerances required for satisfactory operation with the Accu-Chek FastClix device. Use of other lancet drums with the Accu-Chek FastClix device may prevent or impair proper function of the Accu-Chek FastClix device.

Using the Accu-Chek FastClix device indicates your acceptance of the restricted license to use the Accu-Chek FastClix device only with Accu-Chek FastClix lancet drums. If you do not agree to the terms and conditions of the restricted license, you may return, at the place of purchase, the unused Accu-Chek FastClix device for a full refund. If you have any questions, please call the Accu-Chek Customer Care Service Center at 1-800-858-8072.

Except where prohibited by statute, all warranties covering the Accu-Chek FastClix device are voided by use of the Accu-Chek FastClix device with any lancet drums other than Accu-Chek FastClix lancet drums.

Patent Information

U.S. Pat.: <http://www.roche-diagnostics.us/patents>

A

alternate site testing 33
averages 66

B

battery, changing 103
battery type 113
blood glucose results, unusual 34
blood glucose test, performing 25
blood glucose, warnings 29
button, functions 12

C

comfort dial 11, 19
comments, adding to results 30
control results, out-of-range 42
control solution 37
control test, performing 38

D

data transfer 76
 USB 78
 wireless 76

H

high blood glucose 35
hyperglycemia 35
hypoglycemia 35

L

lancet drum, changing 22
lancet drum, inserting 20
lancing device, cleaning and disinfecting 99
lancing device, using 21
logbook 64
low blood glucose 35
low/high data 69

M

maintenance, meter 103
meter, cleaning and disinfecting 96

P

patterns 73
product limitations 113
product safety information 117
product specifications 113

S

safety information 3
settings, meter 43
 beeper 46
 language 61
 last result 59
 patterns 50
 post meal 56
 reminders 53
 target ranges 47
 time/date 45

supplies 119
symbols, blood glucose results 31
symbols, explanation of 119
symbols, meter display 15
symptoms, hypoglycemia/hyperglycemia 35

T

target % data 67
technical information 113
test strips 24

U

use by date 24, 37

W

warranty 121
wireless communication
 auto-send 86
 default device 85
 delete paired device 91
 flight mode 84
 overview 81
 pairing additional device 88
 pairing, first time 82
 sync time 87

Roche Diabetes Care, Inc.
9115 Hague Road
Indianapolis, IN 46256
www.accu-chek.com

ACCU-CHEK, ACCU-CHEK GUIDE, and FASTCLIX are trademarks of Roche.



CONTINUA, the CONTINUA logos and CONTINUA CERTIFIED are trademarks, service marks, or certification marks of the Continua Health Alliance. CONTINUA is a registered trademark in some, but not all countries in which this product is distributed.

The *Bluetooth*® word mark and logos are registered trademarks owned by Bluetooth SIG, Inc. and any use of such marks by Roche is under license.

All other product names and trademarks are the property of their respective owners.

© 2016 Roche Diabetes Care

00100918001-0916



DRAFT